

# Neurorehabilitation and Neural Repair

<http://nnr.sagepub.com/>

---

## First Permanent Implant of Nerve Stimulation Leads Activated by Surface Electrodes, Enabling Hand Grasp and Release: The Stimulus Router Neuroprosthesis

Liu Shi Gan, Einat Ravid, Jan Andrzej Kowalczewski, Jaret Lawrence Olson, Michael Morhart and Arthur Prochazka  
*Neurorehabil Neural Repair* published online 29 September 2011  
DOI: 10.1177/1545968311420443

The online version of this article can be found at:  
<http://nnr.sagepub.com/content/early/2011/09/28/1545968311420443>

---

Published by:



<http://www.sagepublications.com>

On behalf of:



American Society of Neurorehabilitation

Additional services and information for *Neurorehabilitation and Neural Repair* can be found at:

**Email Alerts:** <http://nnr.sagepub.com/cgi/alerts>

**Subscriptions:** <http://nnr.sagepub.com/subscriptions>

**Reprints:** <http://www.sagepub.com/journalsReprints.nav>

**Permissions:** <http://www.sagepub.com/journalsPermissions.nav>

>> [Version of Record](#) - Sep 29, 2011

[What is This?](#)

# First Permanent Implant of Nerve Stimulation Leads Activated by Surface Electrodes, Enabling Hand Grasp and Release: The Stimulus Router Neuroprosthesis

Neurorehabilitation and  
Neural Repair  
XX(X) 1–9  
© The Author(s) 2011  
Reprints and permission: <http://www.sagepub.com/journalsPermissions.nav>  
DOI: 10.1177/1545968311420443  
<http://nnr.sagepub.com>  


Liu Shi Gan, BEng, PhD<sup>1</sup>, Einat Ravid, BEng<sup>1</sup>,  
Jan Andrzej Kowalczewski, BSc, PhD<sup>1</sup>, Jaret Lawrence Olson, MD<sup>1</sup>,  
Michael Morhart, MD<sup>1</sup>, and Arthur Prochazka, BEng, MSc, PhD<sup>1</sup>

## Abstract

**Background.** More than 150 000 neuroprostheses (NPs) have been implanted in people to restore bodily function in a variety of neural disorders. The authors developed a novel NP, the Stimulus Router System (SRS), in which only passive leads are implanted. Each lead picks up a portion of the current delivered through the skin by an external stimulator. **Objective.** The authors report on the first human implant of an SRS. **Methods.** The recipient was a tetraplegic man with bilateral hand paralysis. Three SRS leads were implanted in his right forearm to activate the finger extensors, finger flexors, and thumb flexor. A wristlet containing a surface stimulator and electrodes was used to pass trains of electrical pulses through the skin to each lead. Hand opening and grasp were controlled via a wireless earpiece that sensed small tooth-clicks and transmitted signals to the wristlet. **Results.** The current required to activate the muscles was less than half that required prior to implantation and below perceptual threshold. Maximal grip force and hand opening aperture were both larger using the SRS. The implanted leads have remained functional for 3 years. The recipient reported various tasks of daily life that improved during SRS usage. An electronic counter revealed mean monthly usage of 18.5 hours, equivalent to 55 hours of continuous manual activity. **Conclusions.** This first implant of the SRS indicates that it can be effective and reliable and has potential to provide an alternative to existing NPs.

## Keywords

neuroprosthesis, functional electrical stimulation, spinal cord injury, stroke

Neuroprostheses (NPs) are electronic devices that stimulate nerves to improve bodily function after damage to the central nervous system. They generally comprise hermetically sealed electrical stimulators with leads terminating in epimysial or nerve cuff electrodes.<sup>1,2</sup> Command signals, and in some cases energy, are transmitted at radiofrequencies from an external control unit to the implanted stimulator.


Over the past 45 years, there has been an enormous proliferation of NPs to ameliorate disorders as diverse as pain, deafness, paralysis, epilepsy, Parkinson disease, tremor, spasticity, depression, and bladder dysfunction.<sup>3</sup> We have developed a new type of externally controlled NP, the Stimulus Router System (SRS).<sup>4</sup> The only component implanted in the body is a passive lead (Figure 1A), one end of which, the “pick-up” terminal, is implanted under the skin, and the other end, the “delivery” terminal, is tunneled to a target nerve (Figure 1B). A surface electrode is placed over the implanted pick-up terminal, and a second electrode

is placed nearby. Current pulses are now passed through the skin between the 2 surface electrodes. Some of the current is “captured” by the implanted pick-up terminal and delivered via the lead to the nerve. The basic properties of the SRS were explored in 2 series of animal experiments.<sup>4,5</sup> Results from the studies showed that the SRS is reliable as a long-term NP and can selectively activate deep-lying nerves in a graded manner over the full physiological range. The first human proof of principle of SRS was shown during an acute human intraoperative procedure,<sup>6</sup> where functional muscle contraction was elicited through the SRS at

<sup>1</sup>University of Alberta, Edmonton, Alberta, Canada

## Corresponding Author:

Arthur Prochazka, University of Alberta, 507 HMRC, Edmonton, Alberta T6G 2S2, Canada  
Email: [arthur.prochazka@ualberta.ca](mailto:arthur.prochazka@ualberta.ca)



**Figure 1.** (A) Each Stimulus Router System (SRS) lead consisted of multistranded platinum–iridium wire loosely coiled inside 1.2-mm diameter silicone tubing. At both ends wire protruding from the tube was tightly wound back onto the surface of the tube to form an electrically conductive terminal. The pick-up terminal was 15-mm long, corresponding to a surface area of 0.9 cm<sup>2</sup>. The delivery terminal was 5-mm long, corresponding to a surface area of 0.3 cm<sup>2</sup>. A silicone strip with a structure similar to a tie-wrap was attached with silicone glue to the delivery terminal to form a nerve cuff. Once the cuff was secured on the target nerve, the excess strips of silicone were trimmed. Tie-downs attached to the leads at each end were sutured to the surrounding connective tissues to act as anchors. (B) The delivery terminals of 2 leads were implanted on branches of the anterior interosseous nerve innervating flexor digitorum profundus (FDP) and flexor pollicis longus (FPL) through skin incisions each about 3 cm long. The delivery terminal of a third lead was implanted on the branch of the posterior interosseous nerve innervating extensor digitorum communis (EDC) and abductor pollicis longus (APL). The pick-up ends of the leads were tunneled under the skin from the incisions and anchored subcutaneously about 3 to 4 cm proximal to the posterior and anterior wrist creases. Abbreviations: SRS, Stimulus Router System; FDP, flexor digitorum profundus; FPL, flexor pollicis longus; EDC, extensor digitorum communis; APL, abductor pollicis longus.

surface current amplitudes below sensory perceptual threshold. In a separate pilot human study that evaluated the use of the SRS system for chronic pain treatment,<sup>7</sup> 10 percutaneous SRS leads were implanted temporarily (5 days) in 8 patients with carpal tunnel syndrome and chronic pain. We describe a case report of the first permanent human implant of an SRS for restoration of hand function.

## Methods

### Participant

The study was performed with approval from the Human Ethics Committee of the University of Alberta and the informed consent of the participant. Inclusion criteria included the following: complete or incomplete cervical spinal cord injury (SCI) with tetraplegia for more than 1 year, loss of voluntary grip and release, functional electrical stimulation (FES) elicits enough hand opening and grip to lift a 4-cm-diameter ball, neurologically and medically stable, able to tolerate FES, and sufficient movement of shoulder and elbow.

The participant was a 52-year-old man who suffered a C6/7 SCI in 1998 due to a sports injury. He had voluntary control of shoulder, elbow, and wrist and a weak tenodesis grip. His neurological motor levels are shown in Table 1. Two-point discrimination was measured at the pulp of the thumb. His pinprick and light touch sensations were normal except for diminished touch at C8 level on the right hand side.

### Surgery

The participant chose to have leads implanted in his right, less functional arm. Just prior to surgery, motor and sensory thresholds were determined by applying trains of current pulses between moistened pad electrodes 5 cm in diameter, placed at the optimal stimulation points of the long finger flexors and extensors and a similar sized reference electrode located on the posterior aspect of the forearm 15 cm proximal to the wrist crease. Grip force with and without stimulation was measured with a hand dynamometer (Jamar; Sammons Preston, Nottinghamshire, UK).

**Table 1.** Neurological Motor Levels of the Participant's Upper Limbs

	International Standards for Classification of Spinal Cord Injury	
	Motor Levels	
	Right Side	Left Side
C5	5	5
C6	4	3
C7	1	2
C8	0	0
T1	0	0

Surgery was performed in 2008 under general anesthesia without muscle relaxation. Incisions were made on the anterior and posterior aspects of the forearm (Figure 1B). Three SRS leads (Synapse Biomedical, Cleveland, Ohio) were implanted, 2 on branches of the anterior interosseous nerve innervating flexor digitorum profundus (FDP) and flexor pollicis longus (FPL) and 1 on a branch of the posterior interosseous nerve innervating extensor digitorum communis (EDC) and abductor pollicis longus (APL). The nerve branches were identified using a bipolar hook electrode connected to a nerve stimulator. In each case, about 2 cm of the nerve was dissected free of connective tissues and a custom silicone nerve cuff<sup>8</sup> containing the delivery terminal of an SRS lead was attached to the nerve by inserting the tongue of the cuff into a slit (Figure 1A) and pulling on the tongue until the cuff had an internal diameter approximately 30% greater than that of the nerve. To verify the placement, a pulse train was applied between the pick-up terminal and a  $3.8 \times 7.6$ -cm hydrogel reference electrode attached to the arm proximal to the elbow prior to surgery. The tongue portion of the cuff was trimmed. Wing-like tie-downs were sutured to subcutaneous fascia about 4 cm distal to the cuffs (Figure 1A). The pick-up ends of the leads were then tunneled under the skin to sites proximal to the wrist on the anterior and posterior aspects of the forearm.

After the surgical incisions had been sutured shut,  $3.8 \times 3.8$ -cm hydrogel electrodes were placed on the skin, one over the FPL and FPB pick-up terminals and another over the EDC/APL pick-up terminal. Stimulation was applied in turn between these electrodes and the reference electrode to elicit grasp and hand opening. The surgery took 2 hours, and the participant returned home the next day. He wore a splint on his operated forearm and hand for 10 days.

### Study Design


Six weeks after the operation, the participant was provided with a wireless earpiece and wristlet stimulator (Figure 2A). The stimulator generated trains of pulses that were delivered

through 4 wettable pad electrodes within the wristlet. Three of these were located over the 3 pick-up terminals. The fourth, the reference electrode, was centered 12 cm proximal to the wrist crease on the posterior aspect of the forearm. Positioning of the electrodes within the wristlet was achieved as follows. First, the reference electrode was held in position on the bare forearm with an elastic strap. Next, pulse trains were delivered through the skin in the regions of the pick-up terminals with a 2.5-cm pad electrode to identify optimal stimulation points. Snap connectors with Velcro backing were now positioned at corresponding positions inside the wristlet, so that when the pad electrodes were attached to the snaps and the wristlet was donned, the electrodes pressed against their respective optimal stimulation points. The EDC/APL and reference electrodes were 5 cm in diameter and those over the 2 flexor pick-up terminals (FDP and FPL) were each 2.5 cm in diameter. The stimulator generated pulses of maximally 10 mA and 100 V.

The participant triggered stimulation with small tooth-clicks that were detected by the wireless earpiece, which contained a 3-axis accelerometer (Figure 2A). A microprocessor in the earpiece ran a pattern recognition algorithm that differentiated tooth-clicks from other vibrations such as those associated with speech.<sup>9,10</sup> On detecting a tooth-click, the earpiece transmitted a radiofrequency signal to a receiver/stimulator located in the wristlet. This allowed the participant to advance the stimulator through the sequence hand-open, grasp, relax, with successive tooth-clicks, as shown schematically in Figure 2B. A 3-axis accelerometer in the stimulator provided an alternative means of advancing through the sequence by tapping or bumping the forearm. A ReJoyce workstation was installed in the participant's home, enabling tele-supervised exercise sessions 5 d/wk for 6 weeks.<sup>11</sup> The sessions initially lasted 15 min/d, gradually increasing to 1 h/d after 3 weeks. Tele-supervised exercise sessions were then conducted at monthly intervals for 6 months.

### Outcome Evaluation

Starting from day 10 postoperatively, the following parameters of the SRS implants were measured every week for 2 months, then monthly for 10 months. (a) The amplitudes of current pulses ( $I_{\text{threshold}}$ ) delivered by a  $3.8 \times 3.8$ -cm hydrogel electrode placed over each pick-up terminal and just eliciting movements of the digits. A  $3.8 \times 7.6$ -cm hydrogel electrode served as the reference electrode, placed longitudinally on the dorsal aspect of the forearm, centered 15 cm from the wrist. (b)  $I_{\text{threshold}}$ ,  $I_{\text{max}}$ , maximal grip force, and thumb-finger aperture elicited with the SRS. In each grip force and hand-opening measurement, the stimulation level was increased by the experimenter to the point that no further increase in force or displacement occurred. Grip force was measured with the Jamar dynamometer.



**Figure 2.** (A) The subject controlled hand opening and grasp with a wireless earpiece and a wristlet device. The wristlet held a detachable, rechargeable stimulator. Insulated wires led from the stimulator to Velcro-backed snaps that were positioned within the wristlet so as to overlie the pick-up terminals of the implanted leads when the wristlet was donned. Prior to use, pad electrodes were moistened and attached to the snaps. The wristlet had straps, buckles, and loops that allowed the participant to don and doff it unaided within 2 minutes. The earpiece contained an accelerometer that detected tooth-clicks. The stimulator also contained an accelerometer that detected taps applied to the forearm. (B) Radiofrequency trigger signals transmitted from the earpiece to the stimulator advanced it through a cycle of hand opening, closing, and relaxation. On detection of the first command (tooth-click or tap applied to the stimulator), stimulus pulses were delivered through the EDC/APL lead, causing hand opening. The pulse amplitude ramped up in 0.5 seconds ( $t_{\text{ramp}}$ ) from zero to a level previously set by the participant via the stimulator interface. The second command initiated a ramp-down of EDC/APL stimulation and ramp-ups of stimulation delivered through the FDP and FPL leads, respectively, causing grasp. The therapist or participant could preset a time delay between activation of FDP and FPL ( $t_{\text{delay}}$ ) so that finger flexion occurred before thumb flexion, avoiding curling of the thumb under the index finger. Abbreviations: FDP, flexor digitorum profundus; FPL, flexor pollicis longus; EDC, extensor digitorum communis; APL, abductor pollicis longus.

## Results


### Preoperative Measurements

Without FES, the participant was unable to generate a measurable grip force on the dynamometer. With voluntary flexion at the wrist, he achieved an aperture of hand opening of 4 cm between the tips of the forefinger and thumb.  $I_{\text{threshold}}$  values for flexion and extension movements were 8.5 mA and 7.5 mA, respectively (Figure 3A, preoperative data points). Pulse trains were then applied through 5-cm diameter wetttable pad electrodes strapped to the forearm, one over the optimal stimulation point of FPL/FDP, another over the optimal stimulation point of EDC/APL, and a third, the reference electrode, on the dorsal aspect of the forearm just proximal to the wrist. A maximal grip force of 29.6 N was achieved with a 30/s train of pulses (300 ms, 19 mA) applied to the flexors. A maximal hand-opening aperture of 10 cm was achieved with 300 ms, 17 mA pulses applied to the extensors. The

participant's sensory perceptual thresholds were tested by placing pairs of surface electrodes at 6 different locations on his forearm. The sensory thresholds averaged  $6.1 \pm 1.4$  mA.

### Postoperative Data

After the wounds were sutured at the end of surgery, a self-adhesive surface electrode (3.8 cm  $\times$  3.8 cm; Sterile Post Operative Electrodes, Uni-Patch, Wabasha, Minnesota) was placed on the skin over the 2 pick-up terminals on the anterior forearm to co-activate FPL and FDP. A similar electrode was placed over the pick-up terminal on the posterior forearm to co-activate EDC and APL. Preoperatively, a 3.8 cm  $\times$  7.6-cm reference electrode had been placed proximal to the elbow and left in place for intraoperative and postoperative testing.  $I_{\text{threshold}}$  values for activation of FPL/FDP and EDC/APL were now 1.3 mA and 1.8 mA, respectively (15% and 24% of the preoperative values).




**Figure 3.** Stimulation parameters before surgery with surface stimulation (unconnected data points) and those monitored periodically over 356 days after the 3 SRS leads had been implanted. (A) minimal surface current ( $I_{\text{threshold}}$ ) needed for activation of the target muscles, (B) surface current ( $I_{\text{max}}$ ) needed to elicit maximal grip force and hand opening, (C) maximal grip force measured using the Jamar dynamometer, (D) thumb to index finger (Th-In) and thumb to middle finger (Th-Mid) separations during maximal hand opening elicited by conventional surface stimulation before surgery and with SRS after surgery. Additional data obtained on day 1053 are shown on the far right of each plot. Abbreviations: SRS, Stimulus Router System; FES, functional electrical stimulation; FDP, flexor digitorum profundus; FPL, flexor pollicis longus; EDC, extensor digitorum communis; APL, abductor pollicis longus.

Ten days after surgery,  $I_{\text{threshold}}$  values for FPL, FDP, and EDC/APL were 2.0 mA, 1.8 mA, and 2.8 mA, respectively. With 1 surface electrode overlying the 2 pick-up terminals on the anterior forearm to co-activate FPL and FDP, a maximal grip force of 96.1 N was elicited at a surface current of 3 mA ( $I_{\text{max}}$ ). This was more than double the maximal grip force elicited with surface stimulation prior to surgery, yet the current was 50% of sensory perceptual threshold and 35% of  $I_{\text{threshold}}$  prior to surgery. Muscles directly under the surface electrodes remained inactive.

Within 10 days of implantation, the participant was able to perform handshakes of increasing strength as the amplitudes of pulse trains were gradually increased. He could crush an empty soft drink can and hold a jug containing a

liter of water by the handle. His only conscious sensation during stimulation was a “pulling” of the muscles as they were activated. He did not feel the paresthesiae he had felt during surface FES prior to the implant.

Figure 3 shows measurements of  $I_{\text{threshold}}$ ,  $I_{\text{max}}$ , maximal grip force, and thumb–finger separation obtained up to 35 months (1053 days) postoperatively. As in the chronic animal studies,<sup>12</sup>  $I_{\text{threshold}}$  and  $I_{\text{max}}$  fluctuated and slightly increased for the first 2 months but then stabilized and decreased (Figure 3A and B). Mean maximal grip force 10 days postoperatively was 100 N, more than double that achieved with surface stimulation preoperatively (Figure 3C). However, 7 days later it had declined to 55 N and did not reach 100 N again until day 100.



**Figure 4.** Smooth gradation of grasp force (A) elicited by slowly increasing and decreasing the amplitudes of pulse trains (B, C) delivered through surface electrodes located over the subcutaneous pickup terminals of the implanted FDP and FPL leads. Abbreviations: FDP, flexor digitorum profundus; FPL, flexor pollicis longus.

Two factors may have contributed to the lower forces in the weeks after day 10. First, compression within the nerve cuffs and surgical trauma may have had adverse effects on nerves and muscles. This is addressed in the Discussion. Second, on day 10, to avoid the forearm incisions, the reference electrode was placed proximal to the elbow joint, whereas on all subsequent occasions and also for the surface stimulation tests prior to surgery, it was placed on the forearm. We had shown previously that the placement of the reference electrode with respect to the implanted nerve cuff can affect the force elicited.<sup>5</sup> This raised the question whether the reading on day 10 provided an appropriate baseline. To address this question, 35 months postoperatively, we compared grip forces elicited with the reference electrode placed either on the forearm or above the elbow. For the forearm location, the mean peak force elicited in 3 repetitions was 92.7 N. For the above-elbow location, it was 113.7 N, that is, 20% higher. So the force at day 10 in Figure 3C was probably about 20% higher than it would have been had it been elicited with the reference electrode on the forearm. However, this would still only account for about a third of the drop after day 10.


Figure 3D shows a 14-cm thumb–finger separation during maximal stimulation after 4 months, which was 10 cm

greater than the separation he could achieve with passive tenodesis and 4 cm greater than the separation achieved with surface FES prior to surgery. At the time of submitting this report, the participant had been using the SRS in activities of daily life (ADLs) for 37 months. He had sufficient residual hand function in his left (nonoperated) hand to adjust the stimulation intensity of each channel and the delay between the onset of FDP and FPL stimulation without external assistance. He typically set the intensities to about 4.5 mA for EDL and APL, 3.5 mA for FDP, and 3.0 mA for FPL. He reported sometimes increasing the intensities for tasks requiring more force, or when muscle fatigue had set in.

The ability to grade activation levels with increasing amplitudes of stimulation was verified 34 months after implantation by delivering interleaved pulse trains through electrodes overlying the FPL and FPL pick-up terminals (Figure 4). The grip force thus elicited rose and fell smoothly as the amplitudes of the pulse trains were slowly ramped up and down. Force began increasing when the pulse amplitudes reached 2.5 mA (FDL) and 3.7 mA (FDP) and increased smoothly, albeit nonlinearly, to 90 N at 3.6 mA (FDL) and 5.6 mA (FDP).

In a separate test, the selectivity of stimulation through surface electrodes located over the FDP and FPL leads was explored by delivering pulse trains (10 pulses/s) through a search electrode 2.5 cm in diameter applied to the skin at a series of points along a mediolateral line that passed over the FDP and FPL stimulation points (Figure 5). Stimulation at the FDP point elicited movements in the index finger and thumb, whereas stimulation at the FPL point elicited movements of the middle, ring, and index fingers.  $I_{\text{threshold}}$  increased approximately by the square of distance of the stimulating electrode from the optimal stimulation point, as previously observed in animal experiments (Figure 4 in Gan et al<sup>4</sup>).

Usage was quantified by a counter in the stimulator that recorded the total duration of stimulation regardless of channel (Figure 6). To put the counter readings into context, in a session involving hand function tests lasting about 45 minutes, the counter read 16 minutes. Thus, a counter reading of 1 hour corresponds to manual activity of about 3 hours. The counter readings were recorded monthly up to 10 months after the participant received the stimulator and then at 12 and 15 months. The total duration of stimulation ranged from a few hours to 56 hours per month. Lower usage in the earlier months was attributed to self-reported muscle fatigue and to technical difficulties associated with the stimulator and wireless earpiece that were gradually overcome. Once the reliability of the external hardware had stabilized, the participant became more confident in using the device in ADLs. At month 7, he experienced a shoulder injury while doing unrelated exercises and his SRS usage declined. It resumed the next month. At month 9, stimulator repairs caused a hiatus of about 3 weeks. At month 10, the



**Figure 5.** Selectivity of activation through the 2 flexor pick-up terminals implanted proximal to the wrist crease. Plots of threshold currents ( $I_{\text{threshold}}$ ) just eliciting movements of the index finger; the thumb; and the middle, ring, and small fingers, delivered at a series of locations with a 2.5 cm search electrode moved along a mediolateral line passing over the flexor pick-up terminals. The indifferent electrode was located on the posterior forearm about 10 cm proximal to the wrist crease. Zero on the horizontal axis arbitrarily corresponds to the location with the lowest thresholds of index finger and thumb movements. The separation between the 2 flexor pick-up terminals was 23 mm, with the pick-up terminal for FPL located more medially on the forearm and the pick-up terminal for FDP located more laterally. Abbreviations: FDP, flexor digitorum profundus; FPL, flexor pollicis longus.

participant had a magnetic resonance imaging (MRI) scan for an unrelated condition and subsequently experienced mild muscle soreness in the implanted forearm. He stopped using the device until the soreness resolved 2 weeks later. At that point, stimulation thresholds were similar to those before the MRI scan (see parameters on day 356 postimplantation; Figure 3), indicating that the target nerve was not damaged. However, this does raise the issue of the safety of MRI scans in people with implanted SRS leads. Tests are underway in our laboratory to understand this better.<sup>13</sup>

Figure 6B shows that flexors were activated for longer durations than extensors, namely, 60% to 70% of total stimulation time. This indicates that in ADLs the time spent gripping objects was longer than the time taken to open the hand around objects. The participant reported that his hand opened more quickly and with a larger aperture with continued daily usage of the device. A video clip of the participant controlling hand opening and grip is provided as Supplementary Material.


## Discussion

The main aim of this study was to evaluate the performance of the SRS as a new type of implantable NP. Conventional NPs consist of an implanted stimulator, controlled by radio-frequency transmission of commands and energy from a coil antenna worn outside the body. We chose to test the

SRS in a motor application in the first permanent human implant, primarily because the activation of nerves innervating muscles elicits responses that are unambiguous and relatively easy to measure.

We found that the SRS was able to selectively activate the target muscles in a graded and controlled manner over their full physiological range at surface current levels below those eliciting sensory perception.  $I_{\text{max}}$  and  $I_{\text{threshold}}$  values stabilized after 100 days and remained stable for the rest of the year.

The  $I_{\text{threshold}}$  levels for FDP and FPL in Figure 3A remained remarkably similar throughout the test period, raising the question of whether they really represented separate activation through the FDP and FPL leads or crosstalk between them. Two observations support separate activation. First, the movements generated by applying a 2.5-cm test electrode to the skin separately over the 2 pick-up terminals were quite distinct: flexion of the middle, ring, and small fingers versus flexion of the index and thumb (Figure 5). These movements corresponded to the separate actions of the 2 branches of the anterior interosseus nerve. Second, Figure 5 indicates that an electrode located over the FPL pick-up terminal would have at least double the threshold for activating the middle, ring, and small fingers compared with the one located over the FDP electrode. An electrode located over the FDP pick-up terminal would have about double the threshold for activating the index finger compared with the one located over the FDP electrode.



**Figure 6.** Usage of the Stimulus Router System. Monthly readings were obtained from a digital counter in the stimulator for 15 months after implantation. The readings represent the total stimulation time, not the total time the device was worn. (A) Monthly durations of flexion and extension stimulation. From the 10-month time point onward, only 2 readings were obtained (at 12 and 15 months, marked by asterisks). For consistency, we computed and displayed the mean monthly stimulation durations at these times. The drops in usage at 7 and 9 to 12 months are discussed in the text. (B) Percentage durations of flexion and extension stimulation.

Grip force and hand aperture with SRS stimulation were larger than those achieved with surface FES prior to surgery (Figure 3). Grip force dropped from 100 N at day 10 to 60 N a week later and recovered to 100 N over the next 7 weeks. About a third of the drop can be accounted for by the differences in reference electrode positions (see Results). What could have caused the rest of the drop? The surgical separation of the nerves from their surrounding tissues or a build-up of pressure within the nerve cuffs might have caused some delayed denervation.<sup>14</sup> After day 10, the forearm became slightly swollen and sore, possibly because usage of the system in daily life had commenced too early. The swelling and tenderness gradually resolved over the next few weeks. The drop in force after day 10, and the gradual recovery, coincided with this period. This suggests

that tissue edema may have reduced the ability of the muscles to generate force.

The amount of surface current needed to achieve useful muscle activation with the SRS was much lower than that required with surface FES prior to surgery and generally below sensory perceptual levels. The ratios of  $I_{\max}/I_{\text{threshold}}$  (surface current eliciting maximal contraction/surface current just eliciting a local muscle contraction) for hand closing and opening calculated over 16 measurements were  $1.9 \pm 0.2$  and  $1.3 \pm 0.2$ , respectively. These ratios, which are similar to those observed in the associated animal studies,<sup>4</sup> may serve as guidelines for future studies. For example, if  $I_{\text{threshold}}$  with SRS stimulation is greater than sensory threshold measured prior to surgery, this would be a clear indication that 1 or more components of the SRS are incorrectly located.

The participant has continued to use the SRS on a daily basis. The counter readings ranged from 2 to 56 hours per month. According to our estimate, this corresponds to manual activity of 6 to 168 hours per month, or on average 0.2 to 5.6 hours per day. To our knowledge, this is the first time usage statistics have been obtained with electronic counters in NPs. Normative data on the mean duration per day of hand muscle activity in able-bodied people in different occupations do not exist but would certainly provide additional context.

The counter results suggest that the benefit of using the system outweighed the inconvenience. The participant described improved efficiency (eg, fewer phone handset drops, improved grip of documents, no need to scoop or “sweep objects into his lap, and ability to pick up heavier objects such as a large milk container, open his refrigerator without assistance, and shake hands at social gatherings). He found that the manipulation of small objects was better done without the SRS, as strong activation of finger flexors and extensors reduced dexterity. This FES action is not unique to the SRS.<sup>15</sup> He could don the wristlet in 2 minutes without assistance, but needed help with the earpiece, so on some occasions he used the alternative triggering method of tapping or bumping his forearm.

The SRS provides several potential advantages over existing systems. Greater selectivity is achieved compared with surface FES because current is routed directly to deeplying nerves. Unlike surface FES systems, where the electrodes must be placed over motor points that may be widely distributed and in inconvenient locations, several SRS leads can be routed subcutaneously to a convenient location. The external garments containing the surface electrodes and stimulator are therefore smaller, easier to don and doff, more comfortable, and less conspicuous.

The surgical implantation of the SRS is potentially less invasive and faster than that of NPs that incorporate implantable stimulators, such as the Neurocontrol Freehand and the Finetech STIMuGRIP.<sup>1,2</sup> The STIMuGRIP stimulator is

implanted in the forearm and leaves a noticeable bulge under the skin. External components of the SRS are easy to service and upgrade. On the other hand, the SRS requires surface electrodes to make good electrical contact with the skin, which is less convenient than inductive coupling. The delivery terminals of SRS leads must be placed within 2 or 3 mm of a nerve. In addition, cuffs containing SRS terminals require open surgery. The SRS is limited in the number of individual stimulation channels. NP stimulators controlled by inductive coupling may therefore be preferable in more demanding applications, using, for example, percutaneously placed Bion microstimulators.<sup>16</sup> This study supports the potential utility of the SRS as a new family of neuroprostheses.

### Acknowledgments

We thank Michel Gauthier for designing the stimulator and ear-piece and Kelly Brunton and Kieva Richards for their assistance in clinical evaluations.

### Declaration of Conflicting Interests

The author(s) declared the following potential conflicts of interest with respect to the research, authorship, and/or publication of this article: Arthur Prochazka is a stakeholder in Rehabtronics, Inc, a university spin-off company that holds the intellectual property in the SRS and ReJoyce systems. Jan Andrzej Kowalczewski holds intellectual property in the nerve cuff used in the SRS system.

### Funding

The author(s) disclosed receipt of the following financial support for the research, authorship, and/or publication of this article: We thank the Canadian Institutes of Health Research and the Alberta Heritage Foundation for Medical Research for funding.

### References

1. Peckham PH, Kilgore KL, Keith MW, Bryden AM, Bhadra N, Montague FW. An advanced neuroprosthesis for restoration of hand and upper arm control using an implantable controller. *J Hand Surg Am.* 2002;27:265-276.
2. Spensley J. STIMuGRIP(R); a new hand control implant. *Conf Proc IEEE Eng Med Biol Soc.* 2007;2007:513.
3. Prochazka A. Neuroprosthetics. In: Field-Fote EC, ed. *Spinal Cord Injury Rehabilitation*. Philadelphia, PA: FA Davis; 2009:87-99.
4. Gan LS, Prochazka A, Bornes TD, Denington AA, Chan KM. A new means of transcutaneous coupling for neural prostheses. *IEEE Trans Biomed Eng.* 2007;54:509-517.
5. Gan LS, Prochazka A. Properties of the stimulus router system, a novel neural prosthesis. *IEEE Trans Biomed Eng.* 2010; 57:450-459.
6. Prochazka A, Gan LS, Olson J, Morhart M. First human intra-operative testing of the Stimulus Router System. Paper presented at: 13th Annual International FES Society Conference; September 21-25, 2008; Freiburg, Germany.
7. Deer TR, Levy RM, Rosenfeld EL. Prospective clinical study of a new implantable peripheral nerve stimulation device to treat chronic pain. *Clin J Pain.* 2010;26:359-372.
8. Kowalczewski J, inventor; Angeltear Solutions, Inc (Edmonton, CA), assignee. Adjustable tissue or nerve cuff and method of use. US patent application 20,090,210,042. February 12, 2009.
9. Prochazka A, inventor; Rehabtronics, Inc (Edmonton, CA), assignee. Method and apparatus for controlling a device or process with vibrations generated by tooth clicks. US patent 6,961,623. November 1, 2005.
10. Simpson T, Broughton C, Gauthier MJ, Prochazka A. Tooth-click control of a hands-free computer interface. *IEEE Trans Biomed Eng.* 2008;55:2050-2056.
11. Kowalczewski J, Chong SL, Prochazka A. In-home upper extremity tele-rehabilitation to improve tetraplegic hand function. *Neurorehabil Neural Repair.* In press.
12. Gan LS, Prochazka A. Characterization and optimization of the stimulus router system, a novel neuroprosthesis. Poster presented at: Proceedings of the 12th Annual International FES Society Conference; November 10-14, 2007; Philadelphia, PA.
13. Yeung CJ, Susil RC, Atalar E. RF heating due to conductive wires during MRI depends on the phase distribution of the transmit field. *Magn Reson Med.* 2002;48:1096-1098.
14. Larsen JO, Thomsen M, Haugland M, Sinkjaer T. Degeneration and regeneration in rabbit peripheral nerve with long-term nerve cuff electrode implant: a stereological study of myelinated and unmyelinated axons. *Acta Neuropathol.* 1998; 96:365-378.
15. Popovic D, Stojanovic A, Pjanovic A, et al. Clinical evaluation of the bionic glove. *Arch Phys Med Rehabil.* 1999;80:299-2304.
16. Popovic D, Baker LL, Loeb GE. Recruitment and comfort of BION implanted electrical stimulation: implications for FES applications. *IEEE Trans Neural Syst Rehabil Eng.* 2007; 15:577-586.